

EC Type-Examination Certificate

Number 5568 Issue 2 Extension 1

This is to certify that the product group reference Peha-soft nitrile white comprising the following products

Product Ref.	Description	Classification:	Level	EN 374-2:2003	
Peha-soft nitrile white	Powderfree nitrile exam glove in white	EN388:2003			
		Abrasion resistance	0	Air Leak	PASS
		Blade cut resistance	0	Water Leak	PASS
		Tear resistance	0		
		Puncture resistance	0		
		EN 374-3:2003			
		Ethanol 35%	2	Kohrsolin extra 0.5%	6
		Isopropanol	1	MikrobacTissues	6
		Chlorhexidine 4%	6	Korsolex extra 4%	4
		Glutaraldehyde 4%	6	Korsolex plus 3%	5
		Hydrogen Peroxide 3%	6	Korsolex Med AF 1%	6
		Dismozon pur 1%	6	Dismofix K 1%	6
		Dismozon pur 4%	6	Dismofix B 1%	6
		Kohrsolin FF 0.5%	6		

Technical reports

SATRA: CHM0192956/MP/A, CHM0191795/MP/RS, SPC0197894/1138,
SPC0197894/MB, CHM0197593/MB

has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the Intermediate category (Category II) through:

- i Testing to the following standard: EN 374-1:2003 (Excluding clause 5.3.2)
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:


- 1. Full details of the certification and product are contained in the manufacturer's technical file
- 2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
- 3. This certificate is issued subject to the conditions on the reverse side of this certificate

Signed:



(G Graham)

Signed:



Date:

12th January 2012

On behalf of SATRA